

**U.S. Food and Drug Administration****CENTER FOR DEVICES AND RADIOLOGICAL HEALTH**[FDA Home Page](#) | [CDRH Home Page](#) | [Search](#) | [A-Z Index](#)[Questions?](#)

510 (k) | [Registration & Listing](#) | [Adverse Events](#) | [PMA](#) | [Classification](#) | [CLIA](#)  
[CFR Title 21](#) | [Advisory Committees](#) | [Assembler](#) | [Recalls](#) | [Guidance](#) | [Standards](#)

[New Search](#)[Help](#) | [More About 21CFR](#)

[Code of Federal Regulations]  
 [Title 21, Volume 8]  
 [Revised as of April 1, 2007]  
 [CITE: 21CFR880.5210]



TITLE 21--FOOD AND DRUGS  
 CHAPTER I--FOOD AND DRUG ADMINISTRATION  
 DEPARTMENT OF HEALTH AND HUMAN SERVICES  
 SUBCHAPTER H--MEDICAL DEVICES

[PART 880 -- GENERAL HOSPITAL AND PERSONAL USE DEVICES](#)

Subpart F--General Hospital and Personal Use Therapeutic Devices

Sec. 880.5210 Intravascular catheter securement device.

(a) *Identification.* An intravascular catheter securement device is a device with an adhesive backing that is placed over a needle or catheter and is used to keep the hub of the needle or the catheter flat and securely anchored to the skin.

(b) *Classification.* Class I (general controls). The device is exempt from the premarket notification procedures in subpart E of part 807 of this chapter, subject to the limitations in 880.9.

[45 FR 69682-69737, Oct. 21, 1980, as amended at 59 FR 63011, Dec. 7, 1994; 66 FR 38804, July 25, 2001]

Database Updated April 1, 2007

[CDRH Home Page](#) | [CDRH A-Z Index](#) | [Contact CDRH](#) | [Accessibility](#) | [Disclaimer](#)  
[FDA Home Page](#) | [Search FDA Site](#) | [FDA A-Z Index](#) | [Contact FDA](#) | [HHS Home Page](#)

Center for Devices and Radiological Health / CDRH